
CFDA UPDATE : LIST OF MEDICAL DEVICES EXEMPTED FROM CLINICAL TRIAL REQUIREMENTS



CLASS II DEVICES EXEMPTED FROM CLINICAL TRIAL REQUIREMENTS

Serial number	Product name	Classification code	Product description
1	Preoperative skin preparation	6801	Formed by electric handles host and battery charger. Used in conjunction with a disposable blade, used before surgery to remove the body and the head of hair.
2	Disposable metal clip	6801	Made from metal, polymer or other materials, used for surgical clipping blood vessels or temporary organizations, immediately after removal. Provided sterile, for single use.
3	Hemostatic clip	6801	By clamps, springs and axle components. Made of pure titanium, stainless steel materials. Not provided sterile. Used during temporary occlusion of blood vessels.
4	Expanders	6801	Usually made from stainless steel. Used in spinal surgery to open and expand space.
5	Preoperative skin preparation	6801	Heads (base for ABS Resin, for cutting edge materials such as low carbon steel) and a handle. Cutter head for sterilization products, for one-time use. Used to remove the body and the head of hair, and prepare for medical procedures requiring hair removal.
6	One-time use of dermal curette	6801	Curettes head, handle, can have protective cap. Curettes head made of stainless steel or other non-toxic metallic material, handle made of plastic materials, protective cap; curettes head diameter, shape of the handle can be divided into multiple specifications product sterilized, disposable suitable surgical scraping of necrotic tissue, skin.

Serial number	Product name	Classification code	Product description
7	Head cleaner	6801	Usually made of aluminum oxide layer, a polyester layer, double sided tape and paper gaskets and other components. This product is a sterile, single use. Head cleaner used to clean electro surgical pencil cutter head.
8	Small vessel measuring tape	6802	Angle gauges, callipers marked on the scale. Made of stainless steel material. Non sterile provided, which can be reused. For microsurgery, measure the outside diameter of the small blood vessels.
9	Nerve hooks	6803	Usually made from stainless steel. In spinal surgery for nerve roots, stretching of the soft Dura and other organizations to protect.
10	Brain-Planar positioning feet	6803	By the coronal, sagittal, and intermediate plate is composed of three parts. Single use sterile products. For brain surgery assistant measurement, positioning in the plane.
11	Neurosurgery with a knife cut	6803	Usually formed by the head and stem. Cutter head are made of stainless steel, diamonds. Used in neurosurgical organizations cut off.
12	Meningioma with shears	6803	Usually connected by a pair of blades, blade of the head, and the grip ring. Stainless steel is usually used. For cutting the meninges tissues.
13	Neurosurgery brain pliers	6803	Usually by the Tong beak, stem and stalk, and blade of the Tong beak, and the grip ring. Stainless steel is usually used.Used in forceps, bite in addition to brain tissue, foreign body, increase of biological or remove the tumor.
14	Neurosurgery with tweezer clamps	6803	Usually made up of superimposed in a pair of tail blade.Stainless steel is usually used. When used in brain surgery, clamping of soft tissue or to remove tumors.

Serial number	Product name	Classification code	Product description
15	Neuro-surgical curettes	6803	Usually long and thin design, the proximal handle, a spoon-shaped distal. Stainless steel is usually used. Used to split brain surgery, curettage.
16	Neurosurgery with a drill	6803	Generally made from stainless steel. Can be reused. Neurosurgery brain surgery.
17	Passive equipment for eye surgery	6804 68226823 6866	Passive short-term (Less than 24 hours) Used equipment, used for many years in the market, non-innovative design, new materials, uncoated, no degradation and absorption, without drugs, animal-derived ingredients, generally for the sterilization of single-use products to help complete the eye surgery, check.
18	Transfer Cap / Impression caps	6806	Aseptic transfer caps or die Hat provided by, way of auxiliary instruments for dental implants, made from used materials such as titanium alloys. When used in implants implants and abutments in the mouth where transferred to the working model. Products use raw materials shall comply with the relevant national and industry standards. Product principle, scope, performance and structure identical to the already available products. Exemption does not include the use of new materials, new technology, new designs or products with a new mechanism, new feature.
19	Soft tissue ring cutter	6806	Use with a dental handpiece, from applied materials such as stainless steel, titanium, or to provide sterile. Gum for dental implant surgery. Products use raw materials shall comply with the relevant national and industrial standards, product performance required to meet the applicable parts of the following standards, such as: YY91010, such as the dental Rotary instruments-with the size and basic principles, scope, performance and structure identical to the already available products. Exemption does not include products already on the market

Serial number	Product name	Classification code	Product description
			materials, technology, design, or the mechanism and function of different products.
20	Thoracic and cardiovascular surgery dissector	6807	By stripping the head, stem and stalk and other components. Stainless steel is usually used. For the dissection of the soft tissue, thrombus within a vein or split, and so on, so that the venous flow.
21	Thoracic and cardiovascular surgery with hook	6807	By the head and Shanks and other components. Head with hook. Stainless steel is usually used. Used to pull the heart tissue.
22	Thoracic and cardiovascular surgery with clip	6807	Can be made up of clips, springs and axle. Stainless steel is usually used. For temporary occlusion of blood vessels in the chest cavity surgery or clamp blood vessels to stop bleeding.
23	Thoracic and cardiovascular surgery scissors	6807	Connected by a pair of blades, blade of the head. Stainless steel is usually used. Used to cut or expand coronary arteries of coronary artery cut.
24	Thoracic and cardiovascular surgery forceps	6807	By a pair of grip composite blades, head for straight or curved shape. Stainless steel is usually used. Not provided sterile. Used to hold the tissue and blood vessels.
25	Thoracic and cardiovascular surgery forceps	6807	Connected by a pair of blades, head for the Tong beak. Stainless steel is usually used. Used for holding organs during chest surgery, vascular, free the superior vena cava, aorta.

Serial number	Product name	Classification code	Product description
26	Coronary scissors	6807	Blades connected by a pair of middle head with cutting, shank ring. Stainless steel is usually used. Not provided sterile.Used to cut or expand coronary arteries of coronary artery cut.
27	Aortic retractor	6807	Products by the Department head and handle. Head with hook.Stainless steel is usually used. Not provided sterile. To stretch the aorta.
28	Blood vessel dilators	6807	Usually made up of head, stem and stalk, in General, can be made of stainless steel or titanium, for expanded vascular caliber in cardiothoracic surgery and vascular surgery.Product principle, scope, performance and structure identical to the products already on the market. Exemption does not include: use of new materials, containing special coating products such as polymers, pharmaceuticals, biological products, and has a special structure and scope and other products.
29	Support rods	6809	By the bar and rotating head. For loop Ostomy surgery through supports intestinal mesentery in a fixed position, preventing the return of bowel loops the peritoneal cavity.
30	Mains-powered bone surgery equipment	6810	Network power power bone organization surgery equipment by network power power (also can for network power, and charging battery double power), can by host, and phone (as rig, and milling machine, and mill machine, and saw machine, and screws scored phone, and screws out phone,), and various tool (can matching) and the annex composition (phone and host can one also can separation); equipment can by power, and design, and technology parameter, and annex, and expected uses, different is divided into several model; for surgery Shi on bone organization implementation drill, and milling, and mill, and saw, and file, and screws scored, and Screw removal, and so on. Product

Serial number	Product name	Classification code	Product description
			specifications applicable parts of the following reference can be used, such as:YY/T 0752-2009 electric bone surgical equipment. Note: modifying 2014 Year 12 Bulletin number in the list of products exempted from the first 51 (Mains-powered bone surgery equipment), an increase of " Files, remove screws into the screw " Contents.
31	Canal shovel	6810	Such products are usually made of stainless steel material.For eradication of bone in spine surgery and amended skeletons.
32	Spinal canal files	6810	Such products are usually made of stainless steel material.For eradication of bone in spine surgery and amended skeletons.
33	Spine surgery with gimlet	6810	Head of the products usually made of stainless steel material.For surgery opening in the bones of the spine.
34	Bite of the cervical spine bone forceps	6810	Such products are usually made of stainless steel material.Product performance using applicable parts of the following reference standards, such as:YY/T 1127 bite bone forceps.Biting for cervical spine surgery of bone tissue and repair bones.
35	Bitten by two joints of the cervical spine bone forceps	6810	Such products are usually made of stainless steel material.Product performance using applicable parts of the following reference standards, such as:YY/T 1127 bite bone forceps.Biting for cervical spine surgery of bone tissue and repair bones.

Serial number	Product name	Classification code	Product description
36	Scoliosis correction clamp	6810	Such products are usually made of stainless steel material.Product performance using applicable parts of the following reference standards, such as:YY/T 1127 bite bone forceps.Biting bone tissue for spinal surgery and repair bones.
37	Elbow flat spinous process bone forceps	6810	Such products are usually made of stainless steel material.Product performance using applicable parts of the following reference standards, such as:YY/T 1127 bite bone forceps.Biting bone tissue for orthopedic surgery and repair bones.
38	Gun-shaped bite bone forceps	6810	Such products are usually made of stainless steel material.Product performance using applicable parts of the following reference standards, such as:YY/T1127 bite bone forceps.Biting bone tissue for orthopedic surgery and repair bones.
39	Disc surgery using ring saw	6810	Such products are usually made of stainless steel material.For resection of intervertebral disc surgery, and amended skeletons, drilling or breaking out, and so on.
40	Lamina stripper	6810	Such products are usually made of stainless steel material.Spine surgery stripping laminae attached to the organization.
41	Sounding device for spinal surgery	6810	Such products are often made from stainless steel material, used in spinal surgery cavity after resection of nucleus pulposus of intervertebral tissue depth measurements.
42	Cervical bone chisel	6810	Such products are usually made of stainless steel material.Product performance using applicable parts of the following reference standards, such as:YY/T 91141 orthopaedics, General specification for cutting class. For the cervical spine surgery and bone tissue during the repair.

Serial number	Product name	Classification code	Product description
43	Lamina bone chisel	6810	Such products are usually formed by the head and knife handles, blades are beveled sharp edge. Usually made of stainless steel material. Product performance using applicable parts of the following reference standards, such as:YY/T 91141 orthopaedics, General specification for cutting class. For cutting and trimming during spine surgery Laminectomy.
44	Stripping in front of vertebral body	6810	Such products are usually made of stainless steel material. Split or separation of mucous membrane, the periosteum during spinal surgery, and so on.
45	Cranioplasty material forming mold	6810	Such products are usually formed from materials by modeling positive and negative molds (primarily made of medical silicone rubber material), auxiliary tools of head shape. As cranioplasty material forming dies.
46	Spine surgery-reamers	6810	Such products are usually made of stainless steel material.Used in spinal surgery hole, reaming or remove endplate cartilage layer.
47	Spine surgery with a scraper	6810	Such products are usually made of stainless steel material.Used for spine surgery interbody Fusion will be the nucleus pulposus, fiber ring, such as shaving.
48	Spinal surgical curettes	6810	Such products are usually made of stainless steel material.Spinal posterior surgical curettage has been pounded in the disc, remove the upper and lower end plates and so on.
49	Spinal surgery with bone drilling	6810	Such products are usually composed of a handle and metal parts. Handles are usually made of plastic and metal parts are usually made of stainless steel. Used in spinal surgery to drill a tunnel.
50	And active devices for drill bits	6810	Such products are generally made from stainless steel.Orthopedic surgery for bone drilling. Used in conjunction with the source device.

Serial number	Product name	Classification code	Product description
51	And active devices for saw blades	6810	Such products are generally made from stainless steel. Orthopedic surgery for osteotomy. Used in conjunction with the source device.
52	Curette for vertebroplasty	6810	Such products are usually made up of head and shank. In the proximal handle, spoon-shaped concave tip of the remote for a sharp edge, or it can be double-ended. Generally made from stainless steel. Reusable for scraping lesions, SCAR and granulation tissue within the sinus and bone cavity and potential dead bone or lacunar pathology.
53	Reamed	6810	The products are divided into two categories, a kind of flexible rod, drill bit, needle and needle-feed composition; the other consists of reaming heads, hose assembly parts, filters, sealing head, locking clips and so on. Orthopedic surgery for intramedullary expanding. Used in conjunction with the source device.
54	Posterior margin of processors	6810	The product generally consists of a guide wire position, dilated, high precision drilling, casing and other components. To handle the posterior marginal osteophyte, rigid projections, to dredge the nerve root channels.
55	Single use fiber-loop suture instrument	6810	These products usually made of stitched shell, stitched components and transmission components. For simple lumbar intervertebral disc marrow nuclear excising operation of fiber loop stitch.
56	Vertebroplasty for the auxiliary devices	6810	Metal instruments are usually made of stainless steel material; filling device (Pressure device) Is made of polycarbonate and polyurethane polymers, such as bone cement injection device using polymer or metallic material; by puncture kits and expansion kits, bushings, composed of bone cement injection Kit and filling equipment, not including the vertebral expansion balloon catheter. For percutaneous vertebral body posterior coronary angioplasty / percutaneous vertebroplasty, create a working channel, fill the bone cement, restore the vertebral anatomy.

Serial number	Product name	Classification code	Product description
57	Passive sterilizing orthopedic instruments	6810	Manage categories I Such passive orthopedic surgical instrument, in the form of sterilization package provides management category was upgraded to II Classes, from the clinical trials.
58	And active devices for filing	6810	The product files are generally made from stainless steel. If the handle part, can be made from polymer material. Orthopedic surgery for rasping the bone ends.
59	Bone cement fill Kit	6810	Such products may include bone cement, bone cement pressure, cement mixers, cement filling bone cement, bone cement injection device, injector, propeller with bone cement, bone cement conveying mixed bone cement, bone cement casing, vacuum systems, intramedullary compression device, adapter, such as casing components. Bone cement-gun is divided into reusable and disposable, remaining components are usually disposable. Mixed bone cement for orthopedic surgery, and bone cement injection (into) vertebrae, limb bone parts.
60	The bone sampler	6810	The product formed by the casing and push rod or tube. Metal parts are typically made of stainless steel, handles are usually made using polymer materials. For percutaneous vertebral body posterior coronary angioplasty / bone tissue sampling in percutaneous vertebroplasty.
61	Bone cement pressure plug	6810	Such products typically consists of femoral compressive acetabular pressure and composition, material is polyurethane and Silicon respectively 6810 Rubber, used for intramedullary pressure moulding or acetabulum after injecting bone cement. Disposable, sterile packaging.
62	Posterior spinal distraction apparatus	6810	The working principle of this kind of product is the instrument tip width is greater than height, used to insert a space, then rotate the device 90 So the width orientation from horizontal to vertical, and open space. Usually made of stainless steel material. Can be divided into the straight and curved, there are a variety of sizes. Can be reused. Used in spinal surgery, open space.

Serial number	Product name	Classification code	Product description
63	Spinal surgery using bone-grafting and push	6810	Such products are often made from stainless steel material, used in posterior spinal fusion surgery to implant bones into bone area and compaction.
64	Spinal surgery with bone graft funnels	6810	Such products are often made from stainless steel material, used to import bone graft in posterior spinal fusion surgery bone graft area.
65	Spine surgery using test mode	6810	Such products are usually made from stainless steel, in a posterior spinal fusion surgery to assess the height of intervertebral space, provide a reference for interbody fusion device of choice.
66	Cage device for spine surgery	6810	Such products are often made from stainless steel material, used to grab fusion in the posterior spinal fusion surgery to complete the integration of the implant or taken out.
67	Lamina expanders	6810	Such products are often made from stainless steel material, such as spinal fusion by open Laminectomy surgery to open space.
68	Spine surgery with nerve hooks	6810	Such products are often made from stainless steel material, used to block nerve roots in the spinal surgery, Fusion cage implantation.
69	Spine surgery nerve retractors	6810	Such products are usually made from stainless steel, used in spinal surgery retracted nerve root or soft, the dural SAC.
70	Spine surgery using bone RASP	6810	Such products are often made from stainless steel material, used in spinal surgery end plate, edge of the vertebral bone surface treatment.
71	Disposable cervical forceps	6812	Polymer plastic, by male and female to the pliers and Gill shaft, stretching the cervix. Sterile, single use.

Serial number	Product name	Classification code	Product description
72	Disposable Cervical Dilatation balloon catheter	6812	Products from the catheter, balloon, filling head, catheters, balloons, filling materials for polymer materials. The product is sterile products, for one-time use. The products are mainly used in mechanical expanding cervix.
73	Embryo transfer catheter-assisted	6813	This product is used to supplement and complement an approved device inserted into the uterus of embryo, so that the placement of in vitro fertilization (IVF) The embryo into the uterus cavity. This product made of hard cores, tubes and components, such as the handle.
74	Assisted reproduction micromanipulation tube	6813	In vitro fertilization and micromanipulation tubes made from borosilicate glass and other materials, individually wrapped in polypropylene protection shell. When products are used in assisted reproductive operations micro-operations, according to the purpose can be divided into different follicle cytoplasm sperm injection tube, sperm injection, pipe, tube, andZona drilling needles, polar body biopsy needle, local Anatomy of oocytes tubes, blastomere biopsy needle.
75	Assisted reproductive ducts	6813	By trocar sheath or guiding catheter, catheter tip and tube, inner tube and catheter core components. Intrauterine insemination catheter is designed for vaginal insertion of uterine cavity injected semen and artificial insemination.
76	Oocyte in a systematic	6813	Product generally consists of egg needles, protective cover, handle, break-proof sets, suction tube, vacuum tube, flushing tube and Luer connector structure, egg needle tip can be equipped with the ultrasonic echo tags , Can be developed under ultrasound. Egg design can be single or double lumen of the needle, sterile, product should be provided , for one-time use. This product is suitable for Ultrasound-guided , Transvaginal ovarian follicles within the oocyte.

Serial number	Product name	Classification code	Product description
77	Catheter subcutaneous tunneling tool	6815	This product consists of a rod, handle, obturator and jacket pipes and other components. Products are sterilized, disposable. Used to create the subcutaneous tunnel, help of subcutaneous implantable pump catheter connecting passages.
78	Skin prick	6815	Made from polymethyl methacrylate polymers. Formed by the test head and disk. Test tools for allergen in allergic population.
79	Pen-injectors (without cartridges and needles)	6815	By pen-injectors, Pistons, dose, injection button display window, dose adjustment knob and dose correction button, pen-syringe cap for cartridge warehouse, composition doesn't contain cartridges and needles. Pen-injectors and injection drug use needle and bobbin, completion of the subcutaneous injection of drugs. Pen-injectors for direct contact with all parts and injection sites. Such as products with specific properties or structures cannot be exempted from clinic.Exemption does not include the use of new materials, active ingredients, new technology, new designs or products with a new mechanism, new feature.
80	Skin grafting machine	6816	By making seats, hob components, fastening components, composed of sterile slide, driven by the external source device, used for extended skin grafts in burn treatment.Different specifications of the slides can be extended to achieve skin in different proportions. Working principle: To be extended on the skin in a sterile slide and load placed on the network, and then drive external source device connected to this product (motor), the motor power is delivered to the product and eventually make this product part of hob in a rotating manner to realize the film rolling.

Serial number	Product name	Classification code	Product description
81	Breast implant trial	6816	These products made from silicone rubber, silicone or other materials, to temporarily insert tool helps doctors select the breast implant shapes and sizes. Provided sterile, for single use.
82	Medical head lights	6820	Medical head lamp by lamp, headband, LED Light source, power pod and the charger on. Used in ophthalmic medical areas, worn on the operator's head, providing lighting for clinical examinations.
83	Automatic measurement of noninvasive blood pressure monitors	6820	Automatic measurement of noninvasive blood pressure monitors is mainly composed of host, inflatable tubes, blood pressure cuff, and power supply. This product uses the oscillometric method or other methods of similar, non-invasive measurement of human blood pressure. The product should comply with the YY 0670-2008 Automatic measurement of noninvasive blood pressure meter, the manufacturer should provide reports on the clinical validation of the overall effectiveness of blood pressure measurement.
84	Ocular electrophysiology equipment	6821	Products from the Stimulator, composed of biological signal amplifiers, display device, used to measure ocular electrophysiological signals for doctors to provide reference information. A variety of measurement procedures, ERG, and VEP and the MERG and the MVEP and the EOG .
85	Surface reference electrode	6821	Generally consists of electrodes (adhesive patch, sensors) and connecting lines, Can be classified by shape round, oval, square, and so on. Non-sterile products, for one-time use. Use in the electrophysiology of the heart should be used in conjunction with appropriate electrophysiologic mapping device or catheter, used for catheter location. Product performance indicators adopted applicable parts of the following reference standards, such as: YY/T0196-2005 disposable ECG electrodes.

Serial number	Product name	Classification code	Product description
86	Middle ear Analyzer	6821	By the host, probes, headphones, etc, used for acoustic impedance middle ear and static pressure testing applied to diagnosis of middle ear disorders.
87	To attract / Perfusion system (tube)	6822	This product can be composed of a handle and attract perfusion tubes, connectors, sealing cap, spare tube, seals and other components. This product has been sterilized and disposable. Works with attracting perfusion control unit, used in video-assisted Thoracoscopic and laparoscopic surgery for aspiration and flushing.
88	Laparoscopic trocar	6822	This product can be used by Components, such as puncture cannula and needle core 。 Products are sterilized, disposable. For use with the endoscope, puncture in endoscopic surgery on the human organism and establish abdominal channel. Exemption does not include: use of new materials using non- DEHP plasticizer containing special coating such as functional polymer, pharmaceutical, bio-products, containing active ingredients, new technology, new design or new mechanisms, as well as with special structures, the scope of products.
89	Refractometer	6822	Optometry by the main Department, Ministry of power and supporting Department of Hubei province. Refractometer measuring eye spherical lens diopter diopter, astigmatism, cylinder axis.
90	Test head	6822	Optometry is able to provide ball lenses, cylindrical lenses and prisms and other optical equipment, controlled by manual or automatic way, checked for eye refraction. Product performance using applicable parts of the following reference standards: YY 0674-2008 ophthalmic instruments of Optometry

Serial number	Product name	Classification code	Product description
91	Visual acuity chart projector	6822	Visual acuity chart projector, screen and Foundation by the host and remote control composition , Non-standard sign. The video projected on the screen, used to detect human visual sensitivity, refractive properties and binocular vision.Product performance using applicable parts of the following reference standards:YY 0764-2009ophthalmic instruments - Visual acuity chart projector
92	Stroboscopic light source	6822	Stroboscopic light source formed by the host, lights, microphones, foot switch, powered by mains or batteries, usually used with laryngoscope and monitor. , Accessibility, design, technical parameters can be divided into several types. To provide throat examination with stroboscopic light source.
93	Endoscopic Sinus surgery	6822	Rigid optical endoscopes, generally by the optical imaging system and lighting system. Optical imaging systems, transfer system, the objective lens eyepiece consists of three parts.Lighting system for optical fiber. Are observed by the objective lens into the image, through the relay system would like to like, and transmitted to the eyepiece by the magnification of the eyepiece or camera system for observation. The product provides Imaging for examination and surgery of the paranasal sinuses.
94	Laryngoscope	6822	Rigid optical endoscopes, generally by the optical imaging system and lighting system. Optical imaging systems, transfer system, the objective lens eyepiece consists of three parts.Lighting system for optical fiber. Are observed by the objective lens into the image, through the relay system would like to like, and transmitted to the eyepiece by the magnification of the eyepiece or camera system for observation. The products for examination and surgery of the larynx to provide Imaging.

Serial number	Product name	Classification code	Product description
95	Fibres nasopharyngoscope	6822	Flexible fiberscopes, usually composed of optical imaging systems and lighting systems, optical imaging systems, transfer system, the objective lens eyepiece consists of three parts. Lighting system consists of a mixed arrangement of multiple beams optical fiber, and multiple imaging system consists of regularly arranged light guide fibers. On the objective lens focus imaging the object directly to the fiber surface, each pixel on a fiber array (each fibre dimension) for receiving corresponding locations like light, and transfer the energy to the image issue at the other end of the optical fiber, all pixels in square end all recombinant form a lens focusing light energy, and then by the magnification of the eyepiece or camera system for observation. Used to observation, photography, nose, and throat.
96	Using fiber upper digestive tract	6822	Flexible fiberscopes, usually composed of optical imaging systems and lighting systems, optical imaging systems, transfer system, the objective lens eyepiece consists of three parts. Lighting system consists of a mixed arrangement of multiple beams optical fiber, and multiple imaging system consists of regularly arranged light guide fibers. On the objective lens focus imaging the object directly to the fiber surface, each pixel on a fiber array (each fibre dimension) for receiving corresponding locations like light, and transfer the energy to the image issue at the other end of the optical fiber, all pixels in square end all recombinant form a lens focusing light energy, and then by the magnification of the eyepiece or camera system for observation. Used for upper digestive tract (not including duodenum) were observed, video.

Serial number	Product name	Classification code	Product description
97	Using fiber lower digestive tract	6822	Flexible fiberscopes, usually composed of optical imaging systems and lighting systems, optical imaging systems, transfer system, the objective lens eyepiece consists of three parts. Lighting system consists of a mixed arrangement of multiple beams optical fiber, and multiple imaging system consists of regularly arranged light guide fibers. On the objective lens focus imaging the object directly to the fiber surface, each pixel on a fiber array (each fibre dimension) for receiving corresponding locations like light, and transfer the energy to the image issue at the other end of the optical fiber, all pixels in square end all recombinant form a lens focusing light energy, and then by the magnification of the eyepiece or camera system for observation. Used to observe the lower digestive tract, camera.
98	Passive non-orthopedic surgical instruments endoscopic products	6822	Passive short-term (Less than 24 hours) Using devices, not and has source equipment tie using, market application years, non-innovation function design, non-new material made, no coating, no degradation, and absorption function, not containing drug, and animal source sex components, for auxiliary completed endoscope Xia check, and guided, and out, and sets tied (anastomosis device except), and contrast, and expansion, clip take, and separation, and cutting organization or take stone /Gravel.
99	Passive endoscopic surgery tool	6822	Passive endoscope used for orthopedic surgery. Usually made from metal or polymer material. Materials, design, technical specifications, can be part of, its intended use and is divided into a number of different models and specifications. Use with Endoscope, used in orthopedic surgery, check.
100	Renal Expander / Sheath expansion	6822	Products are generally made from PTFE, polyethylene, PVC, polypropylene, material, press material, design, technical parameters can be divided into several types and specifications, with probes or without probe. Available in sterile form, for one-time use. Surgery for Urology, establishment of percutaneous approaches.

Serial number	Product name	Classification code	Product description
101	Hysteroscopy uterine manipulator	6822	Products for the hysteroscope accessories, Consists of a head, the head seat, block, connecting rods, hooks, handles, and syringe components . Hysteroscopy or House / operation under laparoscope combined with shifting position of the uterus .
102	Endoscope bridge	6822	Products are made of stainless steel, for urological endoscopic accessories. Used to connect the endoscope and the tube lens , providing channels for surgical instruments surgical observation or diagnostic use.
103	Endoscopic obturator	6822	Products made from stainless steel, insert the mirror sheath, with the sheath using the mirror. Urology and Gynecology endoscopy diagnosis and surgical endoscope microscope sheath inserted in front of the sheath closed if the body mirror, avoid damage to body tissues.
104	Endoscope-guided tube sheath	6822	This product can be used by the sheath tube part, main body and sealing cap (if any), and other components. This product is inserted into the puncture or cavity, tube sheath as endoscopic insertion pathways and prevent endoscopic apex of touching the puncture and puncture internal, protective Apex and bending Department of endoscopy.
105	Sinus curettes	6822	This product can be used by Apex, inserts, handles and other parts. Products supplied as sterile, for single use. Endoscope inserted cavities, advance and retreat through the product's operation and curette tip bending operation used to collect tissue or cells within the body cavity.
106	Endoscopic clip pliers and clip removal pliers	6822	This product can be used by clamp, clamp, wheel, handle components, usually of stainless steel. Can be reused. The product as a delivery device used by the specific dimensions of the casing passes ligation or remove pin / clip.

Serial number	Product name	Classification code	Product description
107	Endoscopy with biopsy bags	6822	Products can be used by transportation devices, Nano bag, Ligature wire, distraction forceps and other parts, for Endoscopic surgical instrument accessory products. This product has been sterilized, for one-time use. When used in endoscopic surgery to biopsy sample taken from the operation area.
108	Digestive endoscopy disposable Guidewire	6822	Product Guide body and protection, and other components. Through the alimentary canal endoscope into the cavity, used to guide the catheter to the specified location.
109	Endoscopic forceps	6822	Endoscopic forceps can be used by three-jaw pliers, components such as control handle and sheath. Sterility of this product available for one-time use. Gripping and clamping when organizations use in endoscopic surgery.
110	Disposable and removing the stone balloon	6822	Products by balloon, joints, tubing and other components, including porous cavity with tube, plastic tube. This product has been sterilized and disposable. Used with the endoscope, used to enter the bile duct through duodenal endoscopy, biliary stones.
111	Endoscopy in neurosurgery trims	6822	Neuroendoscopic trims can be used by fine tuning device, fixture fittings, fixed arm and other components. Generally in line with YY/T0294.1 made of stainless steel materials, non-sterile product. For support and micro-adjust neuroendoscope in neurosurgery or puncture.
112	Gastrointestinal mirror spiral tube (no lubricant)	6822	The product consists of an end with spiral, composed of hose with control handle at the top. Also comes with rotating collar at the top of the hose and connection components. Assisted endoscopic insertion and advance to the target site.

Serial number	Product name	Classification code	Product description
113	Pneumoperitoneum needle	6822	From needles, grip, piston, locking connectors, tube core components, such as needles, does not contain III Class I medical devices components. Used for establishing pneumoperitoneum during endoscopic surgery of the abdomen. Basic principles, scope of application, performance and structure identical to the already available products. Exemption does not include: use of new materials, containing special coating products such as polymers, pharmaceuticals, biological products, and has a special structure and the scope of products.
114	Cervical cancer screening Kit	6822	By fluoroscopy rods (excluding reagents), gynecological examination composed of mat towels, cotton swabs. When using fluoroscopy rods inserted into the vagina, exposing the cervix, the discoloration was observed, in vivo screening for cervical cancer using utensils.
115	Endoscopic suturing device	6822	Stainless steel clamps, suture needles, lever, control handle, used with box, applied minimally invasive surgical suturing, suture or continuous suture in the soft tissue. Can be reused.
116	Disposable specimen brush	6822	General pipeline components, brush head Assembly, and plug, and lesion of patients with endoscopic brushing cells and tissues or secretions.
117	Push the non-vascular Stents	6822	For endoscopic push send ureters or other non-vascular stent.
118	LCD Visual acuity chart	6822	LCD Visual acuity chart is mainly composed of liquid crystal displays, hosts, and power lines. For the detection of human visual sensitivity, refractive properties and binocular vision. Product performance using applicable parts of the following reference standards: YY 0764-2009 ophthalmic instruments - Visual acuity chart projector

Serial number	Product name	Classification code	Product description
119	Fiberoptic Bronchoscopy	6822	Flexible fiberscopes, usually composed of optical imaging systems and lighting systems, optical imaging systems, transfer system, the objective lens eyepiece consists of three parts. Lighting system consists of a mixed arrangement of multiple beams optical fiber, and multiple imaging system consists of regularly arranged light guide fibers. On the objective lens focus imaging the object directly to the fiber surface, each pixel on a fiber array (each fibre dimension) for receiving corresponding locations like light, and transfer the energy to the image issue at the other end of the optical fiber, all pixels in square end all recombinant form a lens focusing light energy, and then by the magnification of the eyepiece or camera system for observation. Used for observation of tracheal and bronchial, camera.
120	Ultrasonic pulsed echo imaging device	6823	Using the principle of ultrasonic pulse-echo ultrasonic pulse-echo imaging equipment, complete human organ tissue imaging ultrasound system. Usually by probe (line array, and convex array, and phase control array, and mechanical fan sweep), and ultrasonic launches / received, and signal processing and image displayed, part composition; can by models, and imaging nature (simulation and digital), and probe configuration, and technology parameter, and additional auxiliary function, and expected uses, different is divided into several model; by surface, and by vaginal, and by rectal on heart, and muscle bones, and outside week vascular, parts of ultrasound imaging, Do not include non-routine contact circulating blood and applied in such a manner. Product performance using applicable parts of the following reference standards, such as: GB 10152-2009 b -type ultrasonic diagnosis equipment.

Serial number	Product name	Classification code	Product description
121	Spectrum Doppler ultrasonic diagnostic equipment	6823	Spectrum Doppler ultrasonic diagnostic equipment using principles of Ultrasonic Doppler blood flow direction, speed and other information, after treatment with spectrum-time or flow-time to be displayed, is divided into continuous pulse boduopule and boduopule . Usually probes (Unit, convex array and line array, phased arrays, mechanical sector scan), ultrasonic transmitter / receiver, signal processing and display and other components of the spectrum; according to the model, data acquisition / processing (analog and digital), probe configurations, technical parameters, additional accessibility features, Intended use and is divided into a number of different models; through body surface, transvaginal, transrectal abdominal organs, heart, superficial tissue blood flow to organs and other parts of the test , excluding the application of non-conventional and contact circulating blood .Product performance using applicable parts of the following reference standards, such as:YY 0767-2009 ultrasound color flow imaging systems.
122	Ultrasound color flow imaging equipment	6823	Ultrasound color flow imaging equipment based on ultrasonic pulse-echo imaging using Doppler blood flow imaging and correlation technology to, and superimposed on the color coded information B Models on a gray-scale image be displayed in real time. Usually by probe (convex array, and line array, and phase control array, and mechanical fan sweep), and ultrasonic launches / received circuit, and signal processing and image displayed, part composition; can by models, and imaging nature (simulation and digital), and probe configuration, and technology parameter, and additional auxiliary function, and expected uses, different is divided into several model; by surface, and by vaginal, and by rectal on abdominal organ, and heart, and Superficial tissue blood flow to organs and other parts of the image , not including non-conventional applications and contact circulating blood . Product performance using applicable parts of the following reference standards, such as:GB 10152-2009 b -

Serial number	Product name	Classification code	Product description
			type ultrasonic diagnosis equipment, YY 0767-2009 ultrasound color flow imaging system , and so on.
123	Newborn Blu-ray therapeutic instrument	6824	Newborn by Blu-ray Blu-ray therapeutic instrument lights, control boxes, Rod, chassis and its components. No other hybrid light source. For the treatment of neonatal jaundice. YY 0669-2008 medical electrical equipment-part 2 parts: particular requirements for the safety of infant phototherapy equipment applicable to the product.
124	Electrotherapy products	6826	Products from the host (signal generation and control devices), electrodes, wires, and other accessories. Application frequency of 1kHz-100kHz AC (including sine, pulse wave and modulated wave) treatment and rehabilitation methods. Intended use should reflect the clinical indication and therapeutic effects. Such as: the product of frozen shoulder has anti-inflammatory and analgesic effects. Product performance, clinical therapeutic effect and indications should be registered in the electrotherapy products within the scope of technical guidelines.
125	Cleaning heating Kit	6826	For operation indoor heating of fluids.

Serial number	Product name	Classification code	Product description
126	Medical film digital scanner	6831	Generally by the host (see support, export basket), power adapter, and different technical parameters can be divided into several types, for medical use X Radiographic film digitization.
127	High pressure syringe	6831	Generally by the host (including control / Display unit), injection components, such as the nose, with a remote control device, design, technical parameters, additional accessibility features and is divided into a number of different models. By intravascular injection of contrast to the human body and fluids (normal saline) get the angiogram, in combination with special syringes for CT enhancement and vascular angiography, MR enhancement and angiographic contrast agent injection. Product standard: YY/T 0935-2014 CT angiography injection device-specific technical requirements.
128	Cranial radiotherapy positioning device	6833	U LaTeX pillow base, arm lift, hand lifting screw rod, square frame with a Vernier scale. According to CT or MRI obtained in intracranial tumor location used in radiotherapy positioning for intracranial tumors.
129	Automatic blood cell Analyzer	6840	Automatic blood cell Analyzer by host and annex composition, annex by automatically into sample system, and data calculation and the processing system composition; can by design, and technology parameter, and expected uses, and additional auxiliary function, different is divided into several model; used impedance method, and optical method and flow type cell operation method on blood sample in the of red blood cells, and white cell, and platelet for count, used flow type cell operation method combined dyeing method on network weaving red blood cells for analysis, used than color method on red protein content for detection, And blood cell related parameters can be calculated, including white blood cell classification and three five-classification by blood Analyzer. Product

Serial number	Product name	Classification code	Product description
			performance using applicable parts of the following reference standards, such as:YY/T 0653-2008Hematology Analyzer.
130	Semi-automatic blood cell Analyzer	6840	Semi-automatic blood cell Analyzer and accessories by the host, external dilution, by design, technical parameters, intended use, additional accessibility features and is divided into a number of different models using optical methods and flow cytometry in blood red cells, white cells and platelets count. Colorimetric method is used to detect hemoglobin and blood cell related parameters can be calculated, including white blood cell classification and three five-classification by blood Analyzer. Product performance using applicable parts of the following reference standards, such as:YY/T 0653-2008Hematology Analyzer.
131	Rheology instrument /Viscometer	6840	Rheology instrument / Viscometer can include mechanical, measurement, temperature control device segment, the control segment. Rotary blood rheometer / viscometer test is usually cone - plate, cylinder, double gap cylinder type; capillary blood rheometer / viscometer test usually have contrast method, direct method - method of pressure sensor Press automation fully automatic and semi-automatic blood rheometer / viscometer,according to principle, technical parameters, intended use, additional accessibility and is divided into a number of different models. Blood rheometer / viscometer to rotation and (or) capillary whole blood or plasma viscosity is measured.

Serial number	Product name	Classification code	Product description
132	Erythrocyte sedimentation rate Analyzer	6840	Erythrocyte sedimentation rate Analyzer Detection unit, a control unit, data processing, displaying and printing unit and other components. According to the principles, technical parameters, intended use, additional accessibility and is divided into a number of different models; instrument by vertical deposition method, through the method of optical or infrared scanning, Westergren tubes for measuring vacuum blood collection tube or simulation of erythrocyte sedimentation interface, converting standard Westergren method or method of Wen's erythrocyte sedimentation rate and / or Hematocrit .Product performance using applicable parts of the following reference standards, such as:YY/T 1251-2014 erythrocyte sedimentation rate Analyzer.
133	Glycosylated hemoglobin Analyzer	6840	Glycosylated hemoglobin analyzers by liquid chromatography systems, sample preparation, sample delivery system. By design, technical parameters, intended use, additional accessibility features and is divided into a number of different models; General principles for liquid chromatography, immunoturbidimetric method, affinity chromatography, used in conjunction with glycated hemoglobin column can be used for quantitative detection on glycosylated hemoglobin in the blood.
134	Microbial turbidity meter	6840	Microbial turbidimetric meter varies from light source, detector, calibration tubes, turbidity method for measuring the optical density of the microbial suspension, Maxwell turbidity to determine concentrations of microbial inoculation.

Serial number	Product name	Classification code	Product description
135	Microbial identification, culture and drug sensitivity Analyzer	6840	Microbial identification, culture and drug sensitivity Analyzer usually consists of automatic inoculator, training and monitoring systems, data management systems. Principles of microbial identification using digital identification of microorganisms, microbial suspension will be measured based on colorimetric or fluorescence spectrometry biochemical test results into digital code, by coding the database contents compare, get microbial identification results. Drug sensitivity test based on turbidimetric or fluorescent method broth dilution test instruments detect automatically at certain time intervals measuring turbidity of microbial growth, or determination of the intensity of fluorescent indicator in the medium, or hydrolysis of fluorescence raw material, concluded that the growth of organisms at various drug concentrations under examination the slope, obtained by regression analysis of minimum inhibitory concentrations MIC values. For isolated from human clinical samples (such as body fluids, blood, urine, stool, wound swab) microorganism identification and drug sensitivity analysis in.
136	¹³ C 、 ¹⁴ C Breath Analyzer	6840	¹³ C/ ¹⁴ C Breath Analyzer by a data processing unit, gas detection units, sample gas inlet unit, display unit and air filter unit, in accordance with the principles, design, technical parameters, additional accessibility is divided into a number of different models; for the body ¹³ C/ ¹⁴ C Isotope trace detection.
137	Automated blood culture system	6840	Automatic blood culture instrument by optical / Pressure detection system, temperature control system, data processing system, together with matching blood culture bottles used to detect flasks pH Values, gas pressure, turbidity, change of fluorescein-labeled substrates or metabolites, qualitative detection of microbes in samples such as blood, cerebrospinal fluid, pleural effusion and ascites.

Serial number	Product name	Classification code	Product description
138	Electrolytes, blood gas analysis, biochemical analysis of the detection electrode	6840	Electrode for electrolyte, blood gas analysis, biochemical analysis by single or multiple electrode blocks, based on the principle of ion-selective electrode method and analysis module contains electrolytes, arterial blood gas analysis module, or biochemical analysis of instruments used.
139	Trace element analyzers	6840	Analyzer of trace elements by the host, trace analysis, analysis of workstations, computers and software. Using electrochemical analysis, atomic absorption spectrometry or mass spectrometry, test blood, urine, hair and other trace elements in samples.
140	Automatic bacteria cultivation system	6840	Automatic bacteria cultivation system typically consists of incubation unit, control units, fluorescence detection / Alarm units, display units, etc, may also include air filters, barcode scanning, design, technical parameters, can be expected to use, additional accessibility features and is divided into a number of different models; by creating appropriate environment for quick training, clinical specimens such as blood and body fluids in the aerobic bacteria, anaerobic bacteria, yeasts, fungi and Mycobacteria.
141	Immuno-fluorescence Analyzer	6840	Immuno-fluorescence Analyzer usually consists of samples, reagents, place, microscope slides, computer system components. Derived from the human body for clinical samples for direct / indirect immunofluorescence test, including automated experiment operation, image acquisition and interpretation of the results is not recommended.

Serial number	Product name	Classification code	Product description
142	Pathological scanning image analysis system	6840	Pathological scanning image analysis system in General by the slide loader, biological microscope system control centre (computer, keyboard, mouse, etc), CCD Video camera / Camera, bar code readers, motorized scanning platform and application software (including analysis of local management and remote access management, does not include an analysis of a specific test and automatic diagnostic feature) and so on. As an auxiliary tool for clinical concern cells according to a specific color, density / strength, size / dimension, patterns / patterns and shapes to detect, classify and count to help pathologists for pathology analysis.
143	Slide scanning image analysis system	6840	Slide scanning image analysis system in General by reading instrument, read Imaging workstation, reading microscope slides and slide scanning image system, read the slide workstation, automatic carrier mobile phone and software (including analysis of local management and remote access management, does not include an analysis of a specific test and automatic diagnostic feature) and so on. For clinical use on computer screen and marked lesion in the slides area, help the cytology technicians or pathologists to glass samples for analysis.
144	Blood gas Analyzer	6840	Blood gas Analyzer sample / Processing unit, a test unit, data processing / Display units and accessories (including various types of electrodes PO ₂ Electrodes, pH Electrodes, PCO ₂ Capillary tube assembly electrode, a reference electrode, and so on), by design, technical parameters, intended use, additional accessibility is divided into a number of different models, such as through use of electrochemical principles. Direct measurement of arterial blood for clinical blood pH, carbon dioxide partial pressure (PCO ₂), oxygen partial pressure (PO ₂), blood oxygen saturation (SO ₂), sodium ions (Na ⁺), potassium (k ⁺), calcium (Ca ²⁺), The magnesium ion (Mg ²⁺), chloride (Cl ⁻), glucose (Glu), lactic acid (Lac), blood urea nitrogen (BUN), creatinine (Creat) and total bilirubin (tBil), hemoglobin (tHb) And hematocrit (Hct) or sodium in the

Serial number	Product name	Classification code	Product description
			serum and plasma ion (Na ⁺), potassium (k ⁺), Calcium ions (Ca ²⁺), magnesium ion (Mg ²⁺), chloride (Cl ⁻), glucose (Glu), Lactate (Lac), blood urea nitrogen (BUN), creatinine (Creat) and total bilirubin (tBil) and other parameters.
145	Urine Analyzer	6840	Urine Analyzer consists of micro-optical transmission unit and dedicated software, photography, design, technical parameters, accessibility is divided into a number of different models; for urinary formed elements analysis.
146	Non-self test blood glucose Analyzer	6840	Non-self-test principle detection of analyzers, electrochemical, photochemical capillary whole blood and human body / Or venous whole blood glucose concentrations, usually by the host and accessories, according to principle, design, technical parameters, additional accessibility features and is divided into a number of different models; and the corresponding test strips matching, for medical workers on patients with blood glucose monitoring is not used for diagnosis of diabetes mellitus. Product performance indicators may refer to the following reference standards of the applicable parts, such as:GB/T 19634-2005 in vitro diagnostic test systems test for blood-glucose monitoring system general specifications.

Serial number	Product name	Classification code	Product description
147	Self test with non-ketone body, cholesterol, triglycerides, uric acid, lactic acid, hemoglobin Analyzer	6840	Self test with non-ketone body, cholesterol, triglycerides, uric acid, lactic acid, hemoglobin analyzers, electrochemical, photochemical, detection of capillary whole blood and human body / Or vein full blood in the ketone body, and uric acid, and lactic acid, and cholesterol, and glycerol three ester, and mashing hemoglobin concentration, usually by host and annex composition, can by work principle, and design, and technology parameter, and additional auxiliary function, different is divided into several model; and corresponding strips supporting, for medical workers on patients ketone body, and uric acid, and lactic acid, and cholesterol, and glycerol three ester, and mashing hemoglobin for monitoring, not for related disease diagnosis.
148	Thrombelastograph Analyzer	6840	Thrombelastograph Analyzer Usually composed of mechanical units, measurement units, control units (with temperature control) and show / Print unit . By design, technical parameters, intended use, additional accessibility features and is divided into a number of different models; thrombelastograph Analyzer and reagents used, and blood samples with a viscosity of blood clotting time (r), blood clot strength (MA), the rate of blood clotting (Angle), blood clots forming time (k) and the activated clotting time (ACT), and other indicators of, used to monitor and analyze blood samples together to complement the clinical assessment.
149	Automatic microplate plate washing machine	6840	Automatic microplate plate washer can be used by control circuits, positive pressure, vacuum pump, solenoid valve control part composed, reagent bottle, liquid crystal displays, technical parameters can be divided into several types, for automatic cleaning of flat-bottom, U Bottom, V End of porous plate.

Serial number	Product name	Classification code	Product description
150	Automatic immunological analyser	6840	Usually by sampling Center, and processing center, and waste and supply center, and system control center, composition, through to antigen antibody mutual combined of Immunology reaction for based, using enzyme mark, and La Department elements mark or chemical glow agent mark Antigen antibody, through series level joint zoom reaction, will light signal or signals and analysis real concentration, phase contact, analysis human sample in the of stay measuring of Antigen or antibody. For the analysis of human body fluids, such as tumor marker antigens and antibodies of pathogens, such as quantitative, semi-quantitative or qualitative test.
151	Albumin assay (bromocresol green method)	6840	In a particular pH Conditions (such as: pH4.2) In the buffer, albumin molecule positively charged, and negatively charged bromocresol green (BCG) To form a blue-green compound at specific wavelengths (such as: 628nm) Has absorption peaks.Complex is proportional to the absorbance and albumin concentrations. For the detection of human albumin in the sample content, clinically used to assist in evaluation of liver / kidney and nutritional assessment.
152	Albumin assay (bromocresol purple method)	6840	Bromocresol purple (BCP) Dissolved in Specific pH (For example, pH5.2) Acetate buffer solution, yellow. When it's combined with albumin into the green complexes. At specific wavelengths (such as603nm), the green compound absorbance, concentration of albumin in the sample can be calculated. For the detection of human albumin in the sample content, clinically used to assist in evaluation of liver / kidneyfunction and nutritional assessment.

Serial number	Product name	Classification code	Product description
153	Chloride assay (mercury thiocyanate method)	6840	Role of chloride with mercuric thiocyanate in the sample, build to solutions of mercuric chloride, and release the corresponding amount of thiocyanate ions, the ions combine with reagents of iron ions in orange-red ferric thiocyanate, its colour is proportional to the depth and content of chloride. Determination of chlorine content in human samples, clinic mainly for high blood or low chlorine syndrome diagnosis.
154	Creatinine assay (picric acid method)	6840	Creatinine reacts with alkaline PICRATE, Orange creatinine picric acid complexes, in particular within the solution rate of increase in absorbance is proportional to the creatinine concentration. For detecting content of creatinine in human samples, clinical evaluation of renal function as the main one.
155	Copper detection reagents (PAESAREagent method)	6840	Under acidic conditions, copper ceruloplasmin and albumin in the dissociation, ascorbic acid (reduced form) dissociation of Divalent copper ions into a copper, cuprous ions chromogenic reagent 3,5-DiBr-PAESA Generation blue complexes, by measuring absorbance of blue copper complexes, copper concentration can be calculated. For the detection of copper ion content in human samples, primarily for disorders of copper metabolism in clinical diagnosis.
156	Assay of iron (ferrous doxazosin method)	6840	Combined with transferrin of serum iron in acidic medium transferrin dissociation, were again reduced to ferrous iron reducing agent, and purple red compound ferrous doxazosin generated at specific wavelengths (such as: 562nm) Has the maximum absorption peak at, can be used for colorimetric determination. For the detection of iron content in human samples, mainly used on clinical diagnosis of anemia.

Serial number	Product name	Classification code	Product description
157	Total iron binding capacity assay (FereneMethod)	6840	Under acidic conditions, serum iron and transferrin in isolation, with reducing agents and reagent, produced by blue complexes, this product 600nm Maximum absorption, its intensity is proportional to the content of iron in serum, can be calculated by comparison and calibration of serum iron concentrations. In the presence of excess iron ions in alkaline buffer, serum was not with the transferrin iron-binding all combine with iron ions, remaining iron reducing agents generate blue complex, color after exposure to agent (600nm). By calculating the reduction of iron ions in the buffer to calculate the unsaturated iron binding capacity in serum. Total iron binding capacity for testing human samples, mainly used on clinical diagnosis of anemia.
158	Fructosamine (glycosylated serum protein) detection kit (blue tetrazolium method)	6840	Fructosamine in serum (glycosylated serum protein) is a macromolecule Keto amine compounds. In alkaline conditions can be purple blue tetrazolium reduction a month on behalf of which generates a quantity proportional to the serum fructosamine concentration. By colorimetric method in 540nm(530-550nm), 1-Deoxy- -1- Morpholine calibrator measured fructose react in a month for the production to obtain serum fructosamine concentration. Detects fructosamine in human samples (glycosylated serum protein) content of clinically used for monitoring of blood sugar levels.
159	Phosphorus detection reagents (phosphorus molybdate method)	6840	Inorganic phosphorus in the sample in acidic solution reacts with ammonium molybdate ammonium phosphomolybdate complex directly at specific wavelengths (such as: 340nm 325nm) Measuring absorbance calculate phosphorus content. Used for the determination of phosphorus content in human samples, primarily for disorders of phosphorus metabolism in clinical diagnosis.

Serial number	Product name	Classification code	Product description
160	Total protein assay (biuret method)	6840	In alkaline solution, protein peptide bonds and copper ions combine to form a blue-purple compound. Blue-purple compound at specific wavelengths (such as:546nm) absorbance is proportional to the number of peptide bonds, and so on can calculate the protein content. For the detection of total protein content in human samples, clinically used to assist in evaluation of liver / kidney / bone marrow function.
161	Cerebrospinal fluid /Urine total protein assay (Phenol Red Coke /Pyrogallol red-molybdate method)	6840	In acidic environments, protein and molybdate ion-containing coking Phenol Red, blue compound 600nm Maximum absorption peak.600nm sample protein concentrations and changes in absorbance is proportional to. For detecting cerebrospinal fluid / contents of total protein in urine samples, clinic in the central nervous system / kidney disease diagnosis , secondary heart /Evaluation of thyroid function.
162	Assay of zinc (PAPSReagent method)	6840	Nitro -PAPS And in alkaline solution Zn Reaction, purple compound is generated at specific wavelengths (such as: 570nm) Has a maximum absorption peak. From Cu and Fe ions interference can adjust pH value and add chelate completely eliminated. For the detection of content of zinc ion in human samples, primarily for disorders of zinc metabolism in clinical diagnosis.
163	Calcium assay (azo arsenic III Method)	6840	Azo arsenic III Combines with calcium to form colored compounds at specific wavelengths (such as: 650nm) Has the characteristic absorption peak. The absorbance is proportional to the calcium concentration in the sample. For detecting content of calcium ion in human samples, primarily for disorders of calcium metabolism in clinical diagnosis.
164	Calcium assay (o-cresolphthalein complexone method)	6840	O-cresolphthalein complexone ketone (OCPC) Is a metal complexing indicator, under alkaline conditions, samples of calcium ion with o-cresolphthalein complexone red complexes, absorbance is proportional to the calcium concentration increases with the sample. For detecting content of calcium ion in human samples, primarily for disorders of calcium metabolism in clinical diagnosis.

Serial number	Product name	Classification code	Product description
165	Magnesium assay (xylydyl blue method)	6840	In alkaline solution, magnesium reacts with xylydyl blue a blue-purple complex. Magnesium ion concentrations can be reduced through xylydyl blue absorbance is measured. For detecting content of magnesium ion in human samples, primarily for disorders of magnesium metabolism in clinical diagnosis.
166	Colloidal gold test strip analyzers	6840	Colloidal gold test strip analyzers, images generated by the test on the mechanical positioning unit / Display, data transmission, data processing software / Print unit, calibration accessories;, accessibility, design, technical parameters can be divided into certain models; specific colloidal gold test strip package, used in colloidal gold test strip interpretation of qualitative, semi-quantitative or quantitative analysis.
167	Biochip reading apparatus	6840	Biochip reading apparatus consists of a video microscope, controller, support software (Biochip image analysis system)Composed of three parts, design, technical parameters, can be expected to use, additional accessibility features and is divided into a number of different models through transferring treated DNA chip or chip scanned images, and use special software to identify and analyze it, and supporting the use of gene chip or chip.
168	Nucleic acid hybridization oven	6840	Nucleic acid hybridization instrument varies from sample processing unit, composed of the temperature control unit, mechanical unit, for nucleic acid hybridization by the clinic.
169	Automatic sampling system	6841	Automatic sampling system usually consists of samples, reagents, motion systems, sampling systems, identification systems, composed of constant temperature oscillating systems, washing systems, for clinical samples / The pipetting of reagents, distribution and precise sample / Reagent system, supporting follow-up testing.

Serial number	Product name	Classification code	Product description
170	Peripheral blood sampling for single use	6841	Generally buttons, shells, needle, needles, needle seat springs. The design of products using the contact pressure.Clinic for skin puncture to collect blood samples.
171	Blood temperature controller	6845	Generally control devices, heating devices, monitoring devices, can be heated (electric heating, water heating) is divided into a number of different models, used in conjunction with blood purification device, used for blood circuit of temperature control, maintaining body temperature.
172	Cardiopulmonary bypass equipment - Heat exchange tank	6845	Generally control devices, water recycling equipment, refrigeration, heating devices, monitoring devices, different technical parameters can be divided into several types, and cardio-pulmonary circulation equipment / Artificial heart-lung machine used for temperature regulation of the extracorporeal circulation in heart surgery. Product performance using applicable parts of the following reference standards, such as:GB 12263-2005 artificial heart lung machine heat exchanger tanks.
173	Peritoneal dialysis titanium screw cap fittings	6845	This product is pure titanium materials , Made up of two parts :Titanium joints and screw cap. Titanium fittings for a double seal , Luer lock connectors. For peritoneal dialysis catheter and the short pipe connections. Basic principles, scope of application, performance and structure identical to the already available products. Exemption does not include: use of new materials, as well as with special structures, the scope of products.
174	External short for peritoneal dialysis tube	6845	Includes protective caps, switches, pipe, tube connection port of peritoneal dialysis. For peritoneal dialysis and separation line or loop line connection. Basic principles, scope of application, performance and structure identical to the already available products. Exemption does not include: use of new materials, containing special coating products such as polymers, pharmaceuticals, biological products, and has a special structure and the scope of products.

Serial number	Product name	Classification code	Product description
175	Peritoneal dialysis accessories	6845	Containing silica gel pipe extension pipe, pipe joint, fixed-mount and plate fixation device. Catheter extension tubes and tube joints for single-use sterile products. To assist in peritoneal dialysis. Basic principles, scope of application, performance and structure identical to the already available products. Exemption does not include: use of new materials, with special structures, the scope of products.
176	Peritoneal dialysis catheters	6845	By silicone tubing, fittings, nuts, protection CAP, screw device. This product is used in peritoneal dialysis, the peritoneal catheter for implantation of peritoneal dialysis. Products comply with the mandatory industry standard YY 0030 provisions of the Peritoneal dialysis catheters, and basic principles, scope, performance and structure identical to the already available products. Exemption does not include: use of new materials, containing special coating products such as polymers, pharmaceuticals, biological products, and has a special structure and the scope of products.
177	Outside the peritoneal dialysis takes over	6845	By tube, fittings, locks, fittings, and components, such as protection caps and does not contain iodine solution, such as protection caps III Class I medical devices components. Provided sterile. In peritoneal dialysis with Peritoneal dialysis catheters external parts attached. Basic principles, scope of application, performance and structure identical to the already available products. Exemption does not include: use of new materials, containing special coating products such as polymers, pharmaceuticals, biological products, and has a special structure and the scope of products.
178	Peritoneal dialysate bag heating apparatus	6845	The product formed by the heating Board, power supply and cable connectors. Excluding one-time-use parts. Clinical apply of the product operation on peritoneal dialysis peritoneal dialysate bag before heating. The product is not in contact with dialysate.

Serial number	Product name	Classification code	Product description
179	Electronic pneumatic tourniquet	6854	And cuff formed by the host, used for hemostasis during limb surgery.
180	Fluid heating apparatus	6854	Generally control devices, heating devices, monitoring devices, technology can be divided into several types, and fluid temperature Kit used for operation indoor heating of fluids.
181	Electric check Chair	6854	Electric check Chair usually consists of a base, back panel, Panel, footrest and foot control switch. Used to support patient examination and treatment time.
182	Obstetric table	6854	Beds usually consists of beds, bed frames, electric control system, kit components. For pregnant women before delivery, process, process for later use. The product performance using applicable parts of the following reference:YY0571-2013medical electrical equipment part 2 parts: particular requirements for the safety of electric hospital bed.
183	Enteral pump	6854	Intestinal within nutrition pump main by host, and power adapter, and built-in battery, part composition, can by work principle, and design, and technology parameter, and additional auxiliary function, different is divided into several model; and intestinal within nutrition pump with lost note pipeline supporting using, not contact lost note liquid, for hospital to can regulation of way for patients gastrointestinal Road lost note nutrition liquid, with, only applies Yu intestinal within nutrition lost note. Product performance using applicable parts of the following reference standards, such as:GB 9706.1-2007 medical electrical equipmentpart I: General requirements for safety.
184	Sleeve skin Expander	6854	The cylinder sleeve, spool and cover. Products can be different according to the size and location of the wound surface, combined with traction lines, extending to pull the skin around the wound exhibition, a narrowing margin of distance or close up.

Serial number	Product name	Classification code	Product description
185	Flushing pipes	6854	The product formed by the flush pipe, used for rinsing and Endoscope in the diagnosis and surgical suction.
186	Tourniquet system	6854	Made up of tourniquet, and trachea and sensor used for patient's limbs were temporarily blocking blood supply of limb surgery, surgery provides a bloodless surgical field.
187	Dental handpiece	6855	Dental handpiece is a used dental tools or apparatus needed to pass energy of the handpiece. Products covered in line with YY1045.1 the dental handpiece 1 part: the requirements of the high speed air turbine of high-speed gas turbine, and YY 1045.2 dental handpiece 2 Parts: direct phone and the requirements of the contra-angle handpiece of dental contra-angle handpiece or dental direct mobile phones, as well as the Polish mobile phone.
188	Dental low voltage electric motor	6855	Dental low voltage electric motors are by the host control or as a separate mobile devices. It through the motor cables with the host connected to the electrical power supply and control by the latter. Dental low voltage electric motors are used to provide the driving forces of dental handpieces for dental surgery. Performance indicators reference YY0836-2011 setting.
189	Dental air motors	6855	Dental air motor controlled by the dental unit. Dental air motors are used to provide the driving force of dental handpieces for dental surgery. Performance indicators reference YY0837-2011 setting.

Serial number	Product name	Classification code	Product description
190	Dental implant	6855	Dental implant surgical devices for power-driven bone tissue. According to its technical structure, usually consists of three structures. 1, and built-in type dental planting machine: the structure of products by function control circuit, and motor line cable, and motor, and phone, and foot switch, and water device, composition; 2, and network power power of dental planting machine: the structure of products for independent equipment, by host, and motor line cable, and motor, and phone, and foot switch, and Components, such as water supply equipment; 3 dental implant machine, internal power supply: the structure of the product by the host, battery, charger, mobile phone, water supply and other components. Components are not included in the drill. Dental implant intended use of the machine for dental implant treatment.
191	Root canal preparation	6855	Endodontic root canal surgery to prepare the machine for the power-driven equipment. According to its technical structure, usually consisting of two structures. First by the host, motor cables, motors, mobile phones, foot switches and other components; the second by the host, batteries, chargers, mobile phone and so on. Components are not included in the root canal files. Clinical root canal preparation for root canal treatment, stage of root canal preparation and cleanup. Product contains no apical positioning function (or is called a root-canal length measurement function).
192	Gutta-percha filling apparatus	6855	Gutta-percha filling instrument for root canal surgery of power-driven equipment. Usually consists of a host of products, handle (or handheld), sharp (pushes gutta-percha channel), power supply and other component parts. Gutta-percha filling instrument for clinical use of root canal surgery on root canal filling, intended use is pressurized and softening gutta-percha, and tooth filling to prepared the root canals. Product contains no apical positioning function (or is called a root-canal length measurement function).

Serial number	Product name	Classification code	Product description
193	Root canal post preparation with a Dental drill	6855	Dental Rotary instruments-a, formed by the handle and head-end, bulk material usually apply materials such as stainless steel, titanium. Used during dental root canal post repair, removing part of the root canal filling material, preparation and root canal morphology of root canal post match. Can handle and head the material, type, size, is divided into a number of different models and specifications. Product base material shall comply with the relevant national and industrial standards, performance indicators must meet applicable parts of the following standards, such as:YY91010dental Rotary instruments- size and basic principles, scope, performance and structure identical to the already available products.Exemption does not include products already on the market materials, technology, design, or the mechanism and function of different products.
194	Growing surgical Burs	6855	Dental Rotary instruments-a, formed by the handle and head-end, bulk material usually apply materials such as stainless steel, titanium. Used during dental implant surgery, preparation and use of planting model matches the implant cavity. Can handle and head the material, type, size, is divided into a number of different models and specifications.Product base material shall comply with the relevant national and industrial standards, performance indicators must meet applicable parts of the following standards, such as:YY 91064dental Rotary instruments- steel and Carbide Burs technical conditions,YY91010the dental Rotary instruments Size and basic principles, scope, performance and structure identical to the already available products. Exemption does not include products already on the market materials, technology, design, or the mechanism and function of different products.

Serial number	Product name	Classification code	Product description
195	Implant attachment remove the drill	6855	Dental Rotary instruments-a, formed by the handle and head-end, bulk material usually apply materials such as stainless steel, titanium. Remove the base used for surgical procedures, planting the body damaged screws and clear thread. Can handle and head the material, type, size, is divided into a number of different models and specifications. Product base material shall comply with the relevant national and industrial standards, performance indicators must meet applicable parts of the following standards, such as:YY 91064dental Rotary instruments- steel and Carbide Burs technical conditions,YY91010the dental Rotary instruments Size and basic principles, scope, performance and structure identical to the already available products. Exemption does not include products already on the market materials, technology, design, or the mechanism and function of different products.
196	Root canal instruments	6855	Dental root canal therapy devices, used in conjunction with an active medical device. Formed by the handle and head-end, bulk material usually used materials such as stainless steel, nickel-titanium alloy. For dental root canal treatment of the root canal is expanded, shaping, cleaning, secondary filling.Can handle and head the material, type, size, is divided into a number of different models and specifications. Product base material shall comply with the relevant national and industrial standards, performance indicators must meet applicable parts of the following standards, such as:YY 0803.1dentistry- root canal 1 parts: General requirements and test methods, YY/T 0803.2dentistry- root canal 2 parts: the amplifier , and the basic principles, scope of Composition, performance and structure identical to the already available products. Exemption does not include products already on the market materials, technology, design, or the mechanism and function of different products.

Serial number	Product name	Classification code	Product description
197	Grinding and polishing equipment	6855	Dental Rotary instruments, used in conjunction with dental handpiece. Working part and axial compositions, such as plate, cup-shaped. Of dental tissues in the mouth or grinding and polishing the restoration. According to location and axis of work material, type, size, is divided into a number of different models and specifications. Products use raw materials shall comply with the relevant national and industrial standards, product performance required to meet the applicable parts of the following standards, such as: YY91010, such as the dental Rotary instruments-with the size and basic principles, scope, performance and structure identical to the already available products. Exemption does not include products already on the market materials, technology, design, or the mechanism and function of different products.
198	Root canal post preparation with a Dental drill	6855	Dental Rotary instruments-a, formed by the handle and head-end, bulk material usually apply materials such as stainless steel, titanium. Used during dental root canal post repair, removing part of the root canal filling material, preparation and root canal morphology of root canal post match. Can handle and head the material, type, size, is divided into a number of different models and specifications. Product bulk materials shall comply with GB/T13810 and other related national and industry standards, performance indicators must meet applicable parts of the following standards, such as:YY91010, such as the dental Rotary instruments-with the size and basic principles, scope, performance and structure identical to the already available products. Exemption does not include products already on the market materials, technology, design, or the mechanism and function of different products.

Serial number	Product name	Classification code	Product description
199	Dental machining titanium and titanium alloys	6855	Titanium and titanium alloys for dental machinable metal materials used by the CAD/CAM Manufacture of Crowns, bridges, dental prosthesis, body material and performance indicators should meet the relevant national and industry standards, such as: GB17168 The dentistry Metallic materials for fixed and removable prosthesis, andGB/T 13810 implants of titanium and titanium alloy for part of the material. Exemption does not include products used for implants and accessories.
200	TRANS-	6856	Transport castors for bed frame, bed Board, bed, mattress, side rails, braking systems, hydraulic lift, push handle and control systems, and accessories (such as: restraint, infusion). Apply in hospitals for patient transport or ambulance. Product performance using applicable parts of the following reference standards, such as: YY 0571-2005.
201	Clean bench	6857	Cabinets, blowers, filters, castors, ultraviolet germicidal lamps and lighting, control systems. This product within the work area is clean up ISO 5 class, suitable for medical, medical science experiments and other units and departments need local clean working environment.
202	Boiling electric sterilizers	6857	Mainly by the subject, cover, electrical control consisting of heating pipe and, using electric heating. For boiling sterilization of medical devices.
203	Cleaning and disinfection and sterilization products	6857	Usually cavities, control systems, etc, for cleaning of medical devices (and / Or) disinfection (and / Or) sterilization.For example: cleaning sterilizer, steam sterilizer, autoclave, ethylene oxide sterilization of hydrogen peroxide and formaldehyde sterilizers, ultraviolet sterilizers, sterilizer

Serial number	Product name	Classification code	Product description
204	Liquid nitrogen storage system	6858	Products by liquid nitrogen storage tanks, pipes, liquid level sensors, temperature sensors, Control Panel, display panel and the alarm system. Suitable for cryopreservation of laboratory samples, but should not be used for provisions for assisted reproduction in vitro cell and tissue cryopreservation.
205	Fluoride in caries prevention materials	6863	Performance, structure, composition, purpose, belonging to YY/T 0823 Materials of the dental fluoride varnish fluoride varnish can be completely covered by material and basic principles, scope, performance and composition identical to the already available products. Main ingredients such as fluoride. For the prevention of dental caries. Exemption does not include the use of new materials, active ingredients, new technology, new designs or products with a new mechanism, new feature.
206	Cleaning powders / Sand blasting powder	6863	Used with special dental sandblasting machine , Usually composed of sodium bicarbonate, silica, aluminum hydroxide, calcium carbonate, calcium fluoride, sodium bisulfate, and other components. To get rid of plaque on your teeth, pigments and calculus. Product principle, scope, performance and composition identical to the already available products. Exemption does not include the use of new materials, active ingredients, new technology, new designs or products with a new mechanism, new feature.
207	Silane coupling agent	6863	Are molecules containing two different property groups in organic silicon compounds. Usually triethoxy silane and other components. Used between improved dental prosthesis and prosthesis and dental bonding force between. Product principle, scope, performance and composition identical to the already available products. Exemption does not include the use of new materials, active ingredients, new technology, new designs or products with a new mechanism, new feature.

Serial number	Product name	Classification code	Product description
208	Denture agent	6863	Composed primarily of poly methyl vinyl ether, sodium salt. For the coating to denture surface to make dentures and gums tightly to, maintain the denture. Product principle, scope, performance and composition identical to the already available products. Exemption does not include the use of new materials, active ingredients, new technology, new designs or products with a new mechanism, new feature.
209	Dental attachments	6863	Usually consists of two main components: Yin and Yang, and by structure divided into bolt type, bar-clip, snap type and ball type and other types. For removable partial dentures, overdenture auxiliary retentive restorations. Products use raw materials shall comply with the relevant national and industry standards. Product principle, scope, performance and structure identical to the already available products. Exemption does not include the use of new materials, new technology, new designs or products with a new mechanism, new feature.
210	Try dental paste	6863	Usually composed of glycerol, alumina, silica, and other components. To simulate the permanent repair materials such as composite resin color, color and check the final restoration of tooth color together with him. Product principle, scope, performance and composition identical to the already available products. Exemption does not include the use of new materials, active ingredients, new technology, new designs or products with a new mechanism, new feature.
211	Root canal enlargement solution / Root canal cleaning solution / Root canal lubricants	6863	Usually made of ethylene diamine tetraacetic acid / Ethylene diamine tetraacetic acid (EDTA) , Carboxymethylcellulose, and other components. For root canal preparation and expand lubrication, cleaning, wash the root canal. Product rationale, scope, performance and composition and has listed the same. Exemption does not include products already on the market compared to using new materials, active ingredients, new technology, new designs or products with a new mechanism, new feature.

Serial number	Product name	Classification code	Product description
212	Side buckle	6863	Usually made from stainless steel, titanium and other materials. For the orthodontic treatment, related to orthodontic brackets, orthodontic buccal tube, orthodontic wire and other ancillary uses. Products use raw materials shall comply with the relevant national and industrial standards, basic principles, scope, performance and structure identical to the already available products. Exemption does not include products already on the market were compared using new materials, new technology, new designs or products with a new mechanism, new feature.
213	Pit and fissure sealant	6863	Performance, structure, composition, purpose, belonging to YY 0622-2008 Of the pit and fissure sealant on dental resins full covering of pit and fissure sealant, and basic principles, scope, performance and composition identical to the already available products. Main component is double-triethylene glycol dimethacrylate ester (TEGDMA) and so on. For seal teeth fissure gap and prevent the formation of dental fissure gap tooth decay. Exemption does not include the use of new materials, active ingredients, new technology, new designs or products with a new mechanism, new feature.
214	Orthodontic band	6863	By applied materials such as stainless steels, titanium alloys, materials, physical properties, size, can be divided into several types. For orthodontic treatment orthodontic wire, expansion , elastomer support or other accessories , so to correct the teeth. Products use raw materials shall comply with the relevant national and industry standards. Exemption does not include products already on the market materials, technology, design, or the mechanism and function of different products.
215	Denture fit point indicator	6863	Usually formed by substrate, catalyst and blockers, mainly contain vinyl silicone rubber and silicon dioxide. Coating on surface of dental prostheses, inspection and repair of bad contacts. Product specifications shall comply with the relevant national and industry standards, such as exempting circumstances do not include the use of new materials, active

Serial number	Product name	Classification code	Product description
			ingredients, new technology, new designs or products with a new mechanism, new feature.
216	Disposable infusion paste	6864	Products from the base material, pressure sensitive adhesive composed, absorbing pad, release paper. When used to protect transfusion or transfusion puncture site, fixed duct, and needle handle.
217	Fixing Paster	6864	By backing and adhesive agent composition, single-use sterile products. For fixing catheters or other devices to puncture site, contact the wound or dressing subprime fixed. Meet the food drug armed tubes (2014),177 in its paper " surface duct fixtures " provisions.
218	Wound dressing	6864	Products by liner, adhesive and absorbing pad is comprised of single-use sterile products. Applied to acute nursing care of superficial wounds.
219	Treatment of gauze	6864	X Barium-ray detection lines (film) absorbent cotton gauze, sterile, disposable products. As surgical dressings, with x ray detection function.
220	Sterilized cotton stick/ Ball /	6864	Containing alcohol, iodine, or povidone iodine swab sticks /Ball / Tablets to complete skin disinfection before the injection, infusion.
221	Surgical masks	6864	Products by the face mask and respirator connected together.Products are sterilized, disposable. Suitable clinical worn by medical personnel during invasive operations, for the treatment of patients and medical personnel of implementing invasive operation to provide protection and prevent the spread of blood, body fluids, and spatter.

Serial number	Product name	Classification code	Product description
222	Squeezable-extremity pressure	6864	By pressure on the limbs for periodic inflatable, promoting venous return and secondary prophylaxis of deep vein thrombosis.
223	Medical elastic stockings	6864	Product complies with YY/T 0851-2011 Preventing thrombosis socks or YY/T 0853-2011 Varicose vein compression stockings for medical standards requirement. To promote venous blood flow, prevention of deep vein thrombosis, ease symptoms of varicose veins.
224	Uterine balloon catheter	6866	The product consists of balloons, catheters, Catheter, check valves, joints, Struts, etc, is used to contrast agent injected into the uterus and fallopian tubes, hysterosalpingography, or injection of methylene blue is used in abdominal surgery plastic surgery of the fallopian tube and/ Or tubal anastomosis.
225	Balloon compression device	6866	By charging devices, pressure gauges, connecting pipes and connectors and other components. Filling the balloon and shrinkage. Products material mature mechanism of clear, and have been registered in the products have the same nature.
226	Tube importers	6866	Intubation importer by catheter, hardened bushings, connectors and other components. Products are sterilized, disposable. Glottis in patients with intubation importer applies when showing incomplete , auxiliary insert the tracheal intubation in patients with tracheal gas Insufflation.
227	One-time use of heat and moisture exchangers /Filter	6866	Products can be composed of a casing, filter, heat and mass Exchange components, such as media and interfaces. One-time use, clean packaging or aseptic packaging. This product is suitable for patients with artificial airway , breathing gas for heat and moisture Exchange and / or filter particles. Product recommendation YY/T 0735 anaesthetic and respiratory equipment for humidifying respired gases in heat and moisture exchanger.

Serial number	Product name	Classification code	Product description
228	Endobronchial intubation	6866	Product connectors, tubes, tracheal cavities / Cuff / Indicates whether the balloon, bronchial lumen / Cuff / Indicates whether the balloon, inflatable tube / Components such as valves, guide wire. Products are sterilized, disposable. For main bronchial intubation, allowing selective air / suction, suction and Bronchoscopy for any lung.
229	Double cavity bronchial cannula	6866	Bronchial catheter probe pair of double cavity bronchial cannula and connectors and other components. Products are sterilized, disposable. Double cavity bronchial cannula is placed within the bronchial trunk, can be used for selective air or exhaust gas, suction, single-lung ventilation, or used for Bronchoscopy.
230	Dosing the spray bottle	6866	Products can be made of medical grade PVC, silicone rubber, polypropylene, polyvinyl acetate materials. Divided into the mask and mouth may be made by dosing the spray bottle, medicine cups, trace the inhaler and other components. For aerosol drugs and feed it in the oral or nasal cavity of patients with.
231	Disposable sterile vaginal Speculum	6866	Products supported by the vaginal Speculum tube and core components. Cylinder-shaped vaginal Speculum tube, front located a short slope, enhanced established back-end loop, established ring on one side of the longitudinal crack. Support guide before cone after the cores are cylindrical, tapered front end has a window, after paragraph spacing ring and handle. Vaginal Speculum tube and the support guide block limit after two conical surfaces smooth transition Kit transparent structure, with cone-shaped into the vagina 360degree expansion observed. Vaginal Speculum tube material for medical use PP plastic, supporting core material for medical PCplastic.
232	Cervical sampling device	6866	Used to collect epithelial cells of the cervix, cells, and cell block. Sterile, single use.

Serial number	Product name	Classification code	Product description
233	Airway Exchange catheter	6866	Airway Exchange catheter by conduit, connectors and other components. This product has been sterilized and disposable products. Used for endotracheal intubation in Exchange when required.
234	Ureteral catheter	6866	Ureteral catheter by the drive pipe and the attachment is composed of two parts, the front end can be with or without balloon. Catheter for polymeric materials. Ureteral catheter sterilization products, for one-time use. Suitable for urethral catheterization during surgery, including:(1) blocked ureter to inject contrast agents (2) injected into the ureter lavage fluid (3) drainage of urethra (4) placement, forward or exchange of assistive devices (including Guide).
235	Double cavity catheter	6866	Products by puncture, protectors, pipe clamp, pipe, drop pipe, flow regulators, pipe connectors and other components. Products are sterilized, disposable. For the irrigation of bladder urinary tract, surgical instruments.
236	Ureteral dilatation balloon catheters	6866	Products can be used by components such as balloons, catheters, catheter. Products are sterilized, disposable. For dilatation of ureteral stricture or ureteroscopy operation for ureteral dilatation or stones.
237	Endoscopic abdominal cavity dilator	6866	Products can be made up of tubes, Expander, tube, sealing rings, three one-way valve (seat), import column, check current clamp and helmet and other components. Laparoscopic minimally invasive surgery, outside of the peritoneal cavity and the surgical site among the actions required to establish an operation room.
238	Minimally invasive fascia closure device	6866	Products include firmware and closed thread Threading in two parts. Closed firmware from the body, clutch, proximal wing, bracelets, Windows , on your base plate, guiding hole, outlet, hold points, Control Board. Thread Threading needle and a handle. Products are sterilized,

Serial number	Product name	Classification code	Product description
			disposable. Used to gather in laparoscopic surgery organizations, percutaneous suture and to close surgical incisions.
239	Soft disposable laparoscopic instrument sheath	6866	Soft disposable laparoscopic instrument sheath (soft sheath for short) is a single-incision laparoscopic surgery instruments, can be protected by incision and multi-channel seal body parts. Incisions other than protection of retraction, cut ring base, multi-sealing cap and the base seal mosaic, sealing of valves as single port laparoscopic surgery with endoscopic, pliers, shears and other equipment in and out of the cut sealing operations channel. Products are sterilized, disposable. Product specifications applicable parts of the following reference can be used, such as:YY 0672.1-2008 endoscopic apparatus 1 part laparoscopy trocar.
240	Disposable endoscope sheaths	6866	Disposable endoscope sheaths Compound tubes, plugs, connectors, clamps, flanges, nozzles, spray heads, margin (Lens) , Silicone tube seal components. Products are sterilized, disposable. For upper respiratory tract and vocal in flexible endoscopy or nasopharynx, for the use of the endoscope provides a sterile disposable protective layer.
241	Fibrin glue preparation for single use	6866	Pushing this product can be used by the pharmaceutical, dual-Chamber dispenser, bracket, such as tees, nozzle and needle parts. This product has been sterilized and disposable. After preparation, protein blend plastic component injection should occur wound surface.
242	Disposable incision protector	6866	Disposable incision protector is made of polymer materials.Can be placed by the outer ring, Central and channel components. Products can be divided into higher and higher forms of size is divided into a number of different specifications. Apply to sinus surgery and minimally invasive, extending the incision in the surgical field, protect cuts from damage to reduce wound infection.

Serial number	Product name	Classification code	Product description
243	Medical cleaning head	6866	This product can be washed, flush top parts, and does not contain fluid. The use of polymer materials. This product is provided for sterilization, for one-time use. Used for nasal irrigation in patients with allergic rhinitis, Earwax, middle ear infection, the patient's ear canal washing.
244	Disposable suction tube	6866	This product can be used by the body and joints and other components, tubular forms. This product has been sterilized and disposable. Attract waste for surgery.
245	Drainage bags / Urine bag	6866	This product can be entered by collecting bags, tubes, discharge valves, reflux and drainage pipe joints and other components. Through the piping connected to the drainage catheter in vitro, to form a closed drainage system. Reflux functional solution flow will not return to the body. Material for medical use PVC plastics and polymers such as polypropylene. Sterilized product, for one-time use. Collect secretions and body fluids for patients.
246	Use of disposable negative pressure drainage	6866	This product can be used by the body, spring, connecting components such as pipes, fittings, regulators, with manual vacuum source (vacuum ball), does not include drainage catheter inserted. Products are sterilized, disposable. When used in clinical negative pressure drainage, connected with the drainage catheter inserted to collect drainage.
247	Ostomy bag (including chassis)	6866	This product includes intestinal Pocket (Pocket open or closed pockets) and urine laying pockets. Stoma bag made up mainly of outer bag, chassis and filters. Urine mainly by laying Pocket cavity-bag (anti regurgitation structure), chassis and drain valve. Can be sterile or non-sterile products, for one-time use. Used to process the waste.

Serial number	Product name	Classification code	Product description
248	Ostomy skin Protectant	6864	The product main components for carboxymethyl cellulose sodium, guar gum and xanthan gum, for powdered products. This product is non-sterile product. The products come into contact with intact skin around the stoma, peristomal skin stays dry, reduce the exudation of skin irritation.
249	Ostomy skin protection film	6866	Products for the silicon substrate, can contain a mixture of alkyl polysiloxane. Dries in seconds, without affecting the other Ostomy paste. Used to protect the skin around the stoma.
250	Stoma plug	6866	This product is mainly absorbed by bolts, filters and paste adhesive composition for medical use. Product's main ingredients include carboxymethyl cellulose, carbon foams and polyurethane foams and other materials. This product is non-sterile product. Used for the control of patients with enterostomy stoma fecal excretion, particularly suitable for excretion and stool forming between the stoma twice in defecation collection of Ostomy drain liquid, can be used with the irrigation system.
251	Leak-proof cream	6866	Leak-proof cream is available containing carboxymethyl cellulose sodium, ingredients such as gelatin mixture Pasty.Used to fill uneven skin, preventing leakage and protect the skin when using Pocket; also applies to protection of leakage around the fistula sites skin, bottom of the colostomy and ostomy chassis hole leakage between the skin and fill the skin folds, convex surface of the skin and scarring.
252	Sealing ring () or leak-proof ring ()	6866	This product contains sodium carboxymethyl cellulose sodium, ingredients such as gelatin adhesive cut into strips or rings.Helps protect the skin to prevent its contact with the body fluids. Skin protection can be used as follows: colostomy, Ileostomy and Urology peristomal skin seal.

Serial number	Product name	Classification code	Product description
253	Ostomy chassis	6866	Stoma chassis made up mainly of adhesives, according to there is no snap ring and snap ring diameter sizes, with or without non-woven collar, with or without convex ring, adhesive components, initial openings and is divided into a number of different models and specifications. Need to use with a two-piece Ostomy bag to handle the waste. Product performance using applicable parts of the following reference standards, such as:GB/T 20407.1-2006 Pocket Word;GB/T 20407.2-2006 Pocket-requirements and test methods;GB/T 20407.1-2006 Ostomy bag determination colostomy and Ileostomy bag Pocket odour dispersion.
254	Nasal septum clips	6866	Post nasal cavity operation insert the nasal cavity, nasal septum fixed.
255	Intra-abdominal pressure monitoring Kit	6866	Include abdominal pressure and urine collection devices in two parts. Intra-abdominal pressure monitoring device consists of a connector, the connector CAP, pipeline, pipeline regulator, guide clamp, and filter unit. Accurate urine collection device from the pipe, measuring rooms, collection bags, valves and accessories (hanging / straps / hook). Use connected to the drain port on the catheter for measuring hydrostatic pressure within the bladder, in order to monitor pressure within the abdominal cavity, accurate measurement, and urine.
256	Sea water nasal spray	6866	Sea water nasal spray from the spray bottle, spray nozzles, pumps, and composition. Spray solution for sea water sterile forms; used for nasal irrigation , nasal cavity clean and moist.
257	Total urine	6866	Usually consists of pipe, measuring Chamber, collecting bags, valves and accessories (hanging / Lanyard / Hook). When attached to the catheter to collect and measure the patient urine output per unit of time.

Serial number	Product name	Classification code	Product description
258	Three Chamber feeding tubes	6866	<p>Single use sterile products, for seriously ill patients with enteral feeding, diagnosis, treatment and monitoring of feeding tubes that are designed, suitable for gastric paralysis / Emptying of the stomach, pyloric stenosis patients.</p> <p>This product is a three Chamber feeding tubes, three feeding cavity respectively cavity, attracting cavity, cavity pressure regulation. Feeding tubes can be inserted nasal feeding cavity ends to the jejunum, used for Enteral nutrition liquid fed draw cavity end to stomach, gastric decompression; pressure regulating Chamber end to the stomach, for pressure control in stomach when decompressing.</p>
259	Sterile liquid transfer device	6866	Is formed by the plug, transfer, used in medicine and liquid mixing, transfer.
260	Disposable skin drill	6866	Formed by the ring drill bits and handles, can have protective cap. Sterile, single-use suitable skin drilling.
261	Disposable nasal gastric tube	6866	Usually are made of polymer material, usually made of tubes, joints and other auxiliary tube components (such as wire), etc, available in aseptic form. Usually consists of a nasal insertion, end into the stomach, end into the intestines. Both jejunum nutrition and conveying, filling and pressure within the stomach.
262	Rectal manometry balloon catheter	6866	Balloon into the rectum, urodynamic study of test pressure in the rectum. When using products connected to the urodynamic analysis.

Serial number	Product name	Classification code	Product description
263	Support recycling	6866	Used in endoscopic direct vision or X Offline capture and recycling of ureteral stent.
264	Ostomy paste ring	6866	Main ingredients for copolymer, gelatin and poly propylene glycol and ethylene - Vinyl acetate copolymers. Come into contact with intact skin around the stoma, with pockets /chassis use, non-sterile product. For complete filling peristomal skin of depression / wrinkles.
265	Sinus balloon catheter	6877	Usually consists of balloon catheters, inserting probes and irrigation pipes and other components. This product has been sterilized , disposable. Used in diagnostic and therapeutic procedures , expansion space instruments within the sinus and sinus cavities ; Can also be used to target sinus rinse , To complete the treatment and Diagnostics.
266	Surgical organizations and importers	6877	Bags and organized by the surgical importer components.Surgical body bag by bag and pull the Organization composition, imported by the importer pin and handle are formed. Products are sterilized, disposable. Surgical tissue bag used for operation to remove and / or crushed or separate organizations in the organizational process; assisted introggression importer for endoscopic surgery surgery organized bag.
267	Nose pancreas / Biliary drainage tube	6877	Usually consists of drainage pipe, pipe and is converting tubes and other components. Made from polymer material. Products are sterilized, disposable. Nasal pancreatic / biliary drainage tube is designed for inserting a pancreatic duct endoscopic direct vision / biliary tract, drainage of pancreatic juice / bile.